THIS IS ISSUED IN STRICT CONFIDENCE THAT IT IS NOT USED AS A BASIS FOR MANUFACTURE OR SALE AND THAT IT IS NOT COPIED, REPRINTED OR DISCLOSED TO A THIRD PARTY EITHER WHOLLY OR IN PART WITHOUT THE PRIOR WRITTEN CONSENT OF CAL TEST ELECTRONICS, INC.

REVISIONS								
REV.	DESCRIPTIONS	DATE	APPROVED					
_	CLASS A RELEASED	7/12/99						

- 1. ADAPTER, 6/4mm SPADE LUG WITH 4mm BANANA JACK.
- 2. SPECIFICATIONS:

INSULATOR: POLYACETAL


SPADE BODY: BRASS, NICKEL PLATED

3. RATING:

CURRENT: 36A MAX. RESISTANCE:  $<1 \text{m}\Omega$ .

TEMPERATURE:  $-20^{\circ}$  C TO  $+80^{\circ}$  C.





		UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES	CONTRACT NO.		Cal Test Electronics 22820 Savi Ranch Parkway Yorba Linda, CA 92887-4610						
		TOLERANCES ARE: FRACTIONS DECIMALS									
$-(\oplus)$	<del>) [ -                                  </del>	± 1/32 .XX ± .02 .XXX ±.005 DO NOT SCALE DRAWING	APPROVALS	DATE	TITLE	LE CALL CONDE			ADADTED		
$\downarrow$			DRAWN WDH	7/12/99	6/4mm SPADE L			LUG	.UG ADAPTER		
		MATERIAL	CHECKED				6/4mm SPADE LUG ADAPTER w/4mm JACK				
		FINISH	ENGR.		SIZE	CAGE CODE	4.5	DWG. NO.	Da	55300	REV.
NEXT ASSY	USED ON		DESIGN ACTIVITY		A 43F		45	P		.55500	_
APPLICATION				1	SCALE 1:1				SHEET 1 OF 1		